

ELISA for Quantitative Determination of Rat IgG

Introduction

The CellTrend IgG(rat)-ELISA is designed for the quantitative determination of rat IgG in complex samples (serum or other biological samples).

Principle of the Assay

The determination of rat IgG is carried out as direct sandwich ELISA. An antibody specific for rat IgG has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any IgG present is bound. After washing away any unbound substances, an enzyme-linked antibody is added. Following a wash, a substrate solution is added to the wells and color develops in proportion to the amount of antibody conjugate. The absorption at 450 nm is proportional to the IgG concentration.

Performance Characteristics

Standard curve.

7 standards between 15.6 ng/ml and 1000 ng/ml

Sensitivity:

156 ng/ml (sample dilution 1:10)

Sample materials:

Protein solutions


Intraassay precision (CV):

(n=10)

at 73 ng/ml: 5.8% at 274 ng/ml: 8.9%

For Research use only

Typical Standard Curve



Assay Procedure

Incubation of samples/standards	100 μl/well	1 hr with shaking or 2 hrs without shaking
Wash		
Incubation of detection antibody	100 μl/well	1 hr with shaking or 2 hrs without shaking
Wash		
Substrate incubation	100 µl/well	15 min, room temperature
Add stop solution	100 µl/well	
Read at 450 nm		

Order informations

Product	Catalog number	Price (€)
ELISA for Quantitative Determination of IgG (rat), 1x96 determinat.	52300	490
Contract analysis: Determination of IgG (rat)		please inquire

Related products

Product	Catalog number	Price (€)
ELISA for Quantitative Determination of IgA (rat), 1x96 determinat.	52400	490
ELISA for Quantitative Determination of IgM (rat), 1x96 determinat.	52500	490
ELISA for Quantitative Determination of IgE (rat), 1x96 determinat.	52600	490
Contract analysis		please inquire

CellTrend GmbH, Im Biotechnologiepark 3 (TGZ II), D-14943 Luckenwalde, email: info@celltrend.de
Phone: +49 (0)3371 / 61 99 600, FAX: +49 (0)3371 / 61 99 604

Version: 09/2018